

2015/04/28 現在 118756 Ci  
 2014/12/17 測定 4394 TBq

## 食品照射棟第2照射室 線量率分布

線源中心からの距離(Xcm)

測定点の高さ(22.5cm)

	0	20	40	60	80	100		
5	1.14E+06	1.08E+06	9.45E+05	5.69E+05	2.67E+05	1.34E+05		
10	9.01E+05	8.59E+05	7.41E+05	4.84E+05	2.55E+05	1.34E+05		
20	6.04E+05	5.77E+05	4.98E+05	3.68E+05	2.26E+05	1.34E+05		
30	4.32E+05	4.14E+05	3.61E+05	2.84E+05	1.93E+05	1.27E+05		
50	2.51E+05	2.42E+05	2.17E+05	1.82E+05	1.43E+05	1.10E+05		
90	1.14E+05	1.11E+05	1.04E+05	9.42E+04	8.22E+04	7.01E+04		
130	6.44E+04	6.35E+04	6.09E+04	5.71E+04	5.23E+04	4.71E+04		
170	4.13E+04	4.09E+04	3.97E+04	3.80E+04	3.58E+04	3.33E+04		
210	2.90E+04	2.91E+04	2.82E+04	2.72E+04	2.58E+04	2.44E+04		
250	2.12E+04	2.11E+04	2.09E+04	2.03E+04	1.97E+04	1.90E+04		
270	1.89E+04	1.86E+04	1.84E+04	1.80E+04	1.73E+04	1.66E+04		

距離－線量率曲線

